

August 4, 2009

OUR FILE: 46405-4797

Submission No: 135885

GLOBAL EMC STANDARD TECH. CORP.

No. 3 Pau-Tou-Tsuo Valley
Chia-Pau Tsuen, Lin Kou Hsiang
Taipei, Taiwan 244
R.O.C.

Attention: Tonny Lin

Dear Sir/Madame:

The Bureau has received your application for the renewal of four 3m/10m open area test sites and a 3m alternative test site and the registration of two 3m/10m open area test sites. Be advised that the information received was satisfactory to Industry Canada. The following number(s) is now associated to the site(s) for which registration / renewal was sought (**4797A-1, 4797A-2, 4797A-3, 4797A-4, 4797A-5, 4797A-6 and 4797A-7**). Please reference the appropriate site number in the body of test reports containing measurements performed on the site. In addition, please keep for your records the following information;

- Your primary code is: 4797

- The company number associated to the site(s) located at the above address is: 4797A

| Industry Canada Site Number | Company Site Number | Site Description | Expiry Date (YYYY-MM-DD) |
|------------------------------------|----------------------------|--------------------------|---------------------------------|
| 4797A-1 | A1 | 3m/10m OATS | 2011-08-04 |
| 4797A-2 | A2 | 3m/10m OATS | 2011-08-04 |
| 4797A-3 | B1 | 3m/10m OATS | 2011-08-04 |
| 4797A-4 | B2 | 3m/10m OATS | 2011-08-04 |
| 4797A-5 | A6 | 3m Alternative Test Site | 2011-08-04 |
| 4797A-6 | C1 | 3m/10m OATS | 2011-08-04 |
| 4797A-7 | C2 | 3m/10m OATS | 2011-08-04 |

Furthermore, to obtain or renew a unique site number, the applicant shall demonstrate that the site has been accredited to ANSI C63.4-2003 or later. A scope of accreditation indicating the accreditation by a recognized accreditation body to ANSI C63.4-2003 shall be accepted. Please indicate in a letter the previous assigned site number if applicable and the type of site (example: 3 meter OATS or 3 meter chamber). If the test facility is not accredited to ANSI C63.4-2003 or later, the test facility shall submit test data demonstrating full compliance with the ANSI standard. The Bureau will evaluate the filing to determine if recognition shall be granted.

The frequency for re-validation of the test site and the information that is required to be filed or retained by the testing party shall comply with the requirements established by the accrediting organization. However, in all cases, test site re-validation shall occur on an interval not to exceed two years. There is no fee or form associated with an OATS filing. OATS submissions are encouraged to be submitted electronically to the

Bureau using the following URL;

http://strategis.ic.gc.ca/epic/internet/inceb-bhst.nsf/en/h_tt00052e.html.

If you have any questions, you may contact the Bureau by e-mail at certification.bureau@ic.gc.ca Please reference our file and submission number above for all correspondence.

Yours sincerely,

A handwritten signature in black ink, appearing to be 'JL', written over a horizontal line.

Joshua Laviolette

For: Wireless Laboratory Manager

Certification and Engineering Bureau

3701 Carling Ave., Building 94

P.O. Box 11490, Station "H"

Ottawa, Ontario K2H 8S2

Email: joshua.laviolette@ic.gc.ca

Tel. No. (613) 990-2681

Fax. No. (613) 990-4752